



REF TEK

SYSTEMS INC.

The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Tople Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651

Trimble Navigation Limited
 1600 Tenth Street
 Suite A
 Plano, Texas 75074
 Support phone: 1-888-879-2207
 Support Email: reftek_support@trimble.com
www.trimble.com/infrastructure



REF TEK 147A-01/3/INT/SM P/N:97996
S/N: 15103

Normalized Magnitude-frequency Characteristics

Freq.(Hz)	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
1.00	1.0023	1.0020	1.0014
10.00	1.0000	1.0000	1.0000
20.00	0.9935	0.9945	0.9961
30.00	0.9844	0.9865	0.9904
40.00	0.9744	0.9775	0.9841
50.00	0.9660	0.9693	0.9787
60.00	0.9608	0.9633	0.9751
70.00	0.9608	0.9613	0.9747
80.00	0.9681	0.9644	0.9789
90.00	0.9832	0.9736	0.9878
100.00	1.0072	0.9893	1.0021
120.00	1.0836	1.0413	1.0467
140.00	1.1670	1.1002	1.0962
160.00	1.1402	1.0899	1.0941

Sensitivities

	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
Output Sensitivity (V/g)	2.503	2.501	2.499
Cross-axis Coupling (g/g)	0.52%	0.32%	0.08%