



REF TEK

SYSTEMS INC.

The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Topple Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651

Trimble Navigation Limited
 1600 Tenth Street
 Suite A
 Plano, Texas 75074
 Support phone: 1-888-879-2207
 Support Email: reftek_support@trimble.com
www.trimble.com/infrastructure



REF TEK 147A-01/3/INT/SM P/N:97996
S/N: 15104

Normalized Magnitude-frequency Characteristics

Freq.(Hz)	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
1.00	1.0025	1.0028	1.0028
10.00	1.0000	1.0000	1.0000
20.00	0.9930	0.9923	0.9922
30.00	0.9832	0.9813	0.9813
40.00	0.9726	0.9692	0.9694
50.00	0.9638	0.9587	0.9593
60.00	0.9585	0.9513	0.9528
70.00	0.9589	0.9495	0.9522
80.00	0.9669	0.9547	0.9591
90.00	0.9828	0.9676	0.9740
100.00	1.0077	0.9889	0.9976
120.00	1.0838	1.0563	1.0690
140.00	1.1559	1.1245	1.1323
160.00	1.1075	1.0934	1.0828

Sensitivities

	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
Output Sensitivity (V/g)	2.500	2.499	2.502
Cross-axis Coupling (g/g)	0.92%	0.40%	0.28%