



REF TEK

SYSTEMS INC.

The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Tople Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651

Trimble Navigation Limited
1600 Tenth Street
Suite A
Plano, Texas 75074
Support phone: 1-888-879-2207
Support Email: reftek_support@trimble.com
www.trimble.com/infrastructure



REF TEK 147A-01/3/INT/SM P/N:97996
S/N: 15106

Normalized Magnitude-frequency Characteristics

Freq.(Hz)	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
1.00	1.0025	1.0027	1.0027
10.00	1.0000	1.0000	1.0000
20.00	0.9929	0.9925	0.9924
30.00	0.9829	0.9818	0.9816
40.00	0.9720	0.9701	0.9697
50.00	0.9627	0.9599	0.9593
60.00	0.9568	0.9530	0.9522
70.00	0.9564	0.9516	0.9505
80.00	0.9634	0.9573	0.9560
90.00	0.9784	0.9707	0.9693
100.00	1.0025	0.9927	0.9914
120.00	1.0797	1.0630	1.0628
140.00	1.1647	1.1384	1.1407
160.00	1.1393	1.1164	1.1162

Sensitivities

	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
Output Sensitivity (V/g)	2.501	2.500	2.501
Cross-axis Coupling (g/g)	0.84%	0.24%	0.64%