



REF TEK

SYSTEMS INC.

The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Topple Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651

Trimble Navigation Limited
1600 Tenth Street
Suite A
Plano, Texas 75074
Support phone: 1-888-879-2207
Support Email: reftek_support@trimble.com
www.trimble.com/infrastructure



REF TEK 147A-01/3/INT/SM P/N:97996
S/N: 15123

Normalized Magnitude-frequency Characteristics

Freq.(Hz)	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
1.00	1.0023	1.0030	1.0020
10.00	1.0000	1.0000	1.0000
20.00	0.9936	0.9916	0.9944
30.00	0.9845	0.9798	0.9863
40.00	0.9745	0.9669	0.9776
50.00	0.9658	0.9558	0.9702
60.00	0.9600	0.9483	0.9656
70.00	0.9592	0.9468	0.9659
80.00	0.9651	0.9531	0.9726
90.00	0.9783	0.9677	0.9863
100.00	0.9999	0.9916	1.0079
120.00	1.0701	1.0679	1.0756
140.00	1.1508	1.1475	1.1460
160.00	1.1375	1.1168	1.1167

Sensitivities

	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
Output Sensitivity (V/g)	2.501	2.501	2.501
Cross-axis Coupling (g/g)	0.28%	0.88%	0.16%