



REF TEK

SYSTEMS INC.

The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Tople Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651

Trimble Navigation Limited
1600 Tenth Street
Suite A
Plano, Texas 75074
Support phone: 1-888-879-2207
Support Email: reftek_support@trimble.com
www.trimble.com/infrastructure



REF TEK 147A-01/3/INT/SM P/N:97996
S/N: 15134

Normalized Magnitude-frequency Characteristics

Freq.(Hz)	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
1.00	1.0018	1.0019	1.0023
10.00	1.0000	1.0000	1.0000
20.00	0.9948	0.9945	0.9935
30.00	0.9874	0.9865	0.9840
40.00	0.9791	0.9775	0.9732
50.00	0.9718	0.9694	0.9634
60.00	0.9668	0.9636	0.9559
70.00	0.9660	0.9619	0.9529
80.00	0.9709	0.9660	0.9558
90.00	0.9822	0.9768	0.9657
100.00	1.0011	0.9954	0.9836
120.00	1.0636	1.0602	1.0465
140.00	1.1375	1.1446	1.1270
160.00	1.1228	1.1444	1.1236

Sensitivities

	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
Output Sensitivity (V/g)	2.501	2.501	2.501
Cross-axis Coupling (g/g)	0.36%	0.24%	0.60%