



REF TEK

SYSTEMS INC.

The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Topple Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651

Trimble Navigation Limited
 1600 Tenth Street
 Suite A
 Plano, Texas 75074
 Support phone: 1-888-879-2207
 Support Email: reftek_support@trimble.com
www.trimble.com/infrastructure



REF TEK 147A-01/3/INT/SM P/N:97996
S/N: 15135

Normalized Magnitude-frequency Characteristics

Freq.(Hz)	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
1.00	1.0025	1.0026	1.0028
10.00	1.0000	1.0000	1.0000
20.00	0.9929	0.9926	0.9921
30.00	0.9828	0.9821	0.9810
40.00	0.9717	0.9706	0.9688
50.00	0.9620	0.9607	0.9584
60.00	0.9553	0.9541	0.9515
70.00	0.9538	0.9530	0.9504
80.00	0.9589	0.9591	0.9568
90.00	0.9712	0.9729	0.9713
100.00	0.9914	0.9954	0.9947
120.00	1.0549	1.0662	1.0676
140.00	1.1183	1.1383	1.1396
160.00	1.0878	1.1047	1.1045

Sensitivities

	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
Output Sensitivity (V/g)	2.501	2.501	2.501
Cross-axis Coupling (g/g)	0.24%	0.44%	0.16%