



REF TEK

SYSTEMS INC.

The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Topples Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651

Trimble Navigation Limited
 1600 Tenth Street
 Suite A
 Plano, Texas 75074
 Support phone: 1-888-879-2207
 Support Email: reftek_support@trimble.com
www.trimble.com/infrastructure



REF TEK 147A-01/3/INT/SM P/N:97996
S/N: 15141

Normalized Magnitude-frequency Characteristics

Freq.(Hz)	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
1.00	1.0018	1.0024	1.0029
10.00	1.0000	1.0000	1.0000
20.00	0.9949	0.9933	0.9920
30.00	0.9876	0.9837	0.9806
40.00	0.9794	0.9729	0.9682
50.00	0.9722	0.9633	0.9575
60.00	0.9674	0.9563	0.9503
70.00	0.9667	0.9541	0.9490
80.00	0.9717	0.9582	0.9554
90.00	0.9831	0.9693	0.9700
100.00	1.0019	0.9885	0.9939
120.00	1.0645	1.0526	1.0691
140.00	1.1430	1.1283	1.1431
160.00	1.1472	1.1179	1.0993

Sensitivities

	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
Output Sensitivity (V/g)	2.501	2.498	2.499
Cross-axis Coupling (g/g)	0.16%	0.72%	0.16%