



REF TEK

SYSTEMS INC.

The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Topple Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651

Trimble Inc.
5217 Tennyson Parkway
Suite 400
Plano, TX 75024
Support Phone: 1-888-879-2207
Support Email: MonSol_Support@trimble.com
www.trimble.com/monitoring



REF TEK 147A-01/3/INT/SM

S/N: 16034

Normalized Magnitude-frequency Characteristics

Freq.(Hz)	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
1.00	1.0044	1.0035	1.0053
10.00	1.0000	1.0000	1.0000
20.00	0.9886	0.9904	0.9872
30.00	0.9743	0.9771	0.9730
40.00	0.9611	0.9631	0.9614
50.00	0.9523	0.9516	0.9542
60.00	0.9491	0.9443	0.9515
70.00	0.9522	0.9435	0.9536
80.00	0.9623	0.9503	0.9611
90.00	0.9780	0.9648	0.9735
100.00	0.9993	0.9871	0.9912
120.00	1.0551	1.0523	1.0434
140.00	1.1039	1.1138	1.1005
160.00	1.0812	1.0886	1.0920

Sensitivities

	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
Output Sensitivity (V/g)	2.502	2.500	2.502
Cross-axis Coupling (g/g)	0.40%	0.52%	0.68%