



REF TEK

SYSTEMS INC.

The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Topple Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651

Trimble Inc.
5217 Tennyson Parkway
Suite 400
Plano, TX 75024
Support Phone: 1-888-879-2207
Support Email: MonSol_Support@trimble.com
www.trimble.com/monitoring



REF TEK 147A-01/3/INT/SM

S/N: 16035

Normalized Magnitude-frequency Characteristics

Freq.(Hz)	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
1.00	1.0046	1.0038	1.0041
10.00	1.0000	1.0000	1.0000
20.00	0.9882	0.9896	0.9892
30.00	0.9739	0.9758	0.9758
40.00	0.9613	0.9620	0.9633
50.00	0.9533	0.9513	0.9549
60.00	0.9509	0.9456	0.9517
70.00	0.9546	0.9463	0.9543
80.00	0.9652	0.9543	0.9632
90.00	0.9816	0.9691	0.9773
100.00	1.0038	0.9906	0.9966
120.00	1.0638	1.0505	1.0483
140.00	1.1172	1.1050	1.0981
160.00	1.0858	1.0802	1.0884

Sensitivities

	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
Output Sensitivity (V/g)	2.501	2.500	2.502
Cross-axis Coupling (g/g)	0.28%	0.72%	0.88%