



REF TEK

SYSTEMS INC.

The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Tople Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651

Trimble Inc.
5217 Tennyson Parkway
Suite 400
Plano, TX 75024
Support Phone: 1-888-879-2207
Support Email: MonSol_Support@trimble.com
www.trimble.com/monitoring



REF TEK 147A-01/3/INT/SM

S/N: 16037

Normalized Magnitude-frequency Characteristics

Freq.(Hz)	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
1.00	1.0026	1.0026	1.0021
10.00	1.0000	1.0000	1.0000
20.00	0.9927	0.9927	0.9941
30.00	0.9823	0.9824	0.9856
40.00	0.9708	0.9710	0.9761
50.00	0.9607	0.9611	0.9679
60.00	0.9537	0.9544	0.9622
70.00	0.9519	0.9529	0.9610
80.00	0.9569	0.9585	0.9659
90.00	0.9696	0.9718	0.9774
100.00	0.9910	0.9938	0.9966
120.00	1.0611	1.0655	1.0593
140.00	1.1405	1.1475	1.1341
160.00	1.1196	1.1334	1.1317

Sensitivities

	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
Output Sensitivity (V/g)	2.501	2.499	2.500
Cross-axis Coupling (g/g)	0.04%	0.08%	0.08%