



REF TEK

SYSTEMS INC.

The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Topple Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651

Trimble Inc.
5217 Tennyson Parkway
Suite 400
Plano, TX 75024
Support Phone: 1-888-879-2207
Support Email: MonSol_Support@trimble.com
www.trimble.com/monitoring



REF TEK 147A-01/3/INT/SM

S/N: 16039

Normalized Magnitude-frequency Characteristics

Freq.(Hz)	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
1.00	1.0025	1.0024	1.0021
10.00	1.0000	1.0000	1.0000
20.00	0.9931	0.9932	0.9942
30.00	0.9833	0.9834	0.9858
40.00	0.9725	0.9727	0.9765
50.00	0.9631	0.9633	0.9683
60.00	0.9567	0.9570	0.9626
70.00	0.9555	0.9558	0.9614
80.00	0.9611	0.9616	0.9662
90.00	0.9741	0.9749	0.9778
100.00	0.9956	0.9971	0.9970
120.00	1.0651	1.0701	1.0612
140.00	1.1421	1.1577	1.1409
160.00	1.1209	1.1487	1.1439

Sensitivities

	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
Output Sensitivity (V/g)	2.501	2.501	2.499
Cross-axis Coupling (g/g)	0.16%	0.52%	0.76%