



REF TEK

SYSTEMS INC.

The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Topple Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651

Trimble Inc.
5217 Tennyson Parkway
Suite 400
Plano, TX 75024
Support Phone: 1-888-879-2207
Support Email: MonSol_Support@trimble.com
www.trimble.com/monitoring



REF TEK 147A-01/3/INT/SM

S/N: 16042

Normalized Magnitude-frequency Characteristics

Freq.(Hz)	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
1.00	1.0033	1.0022	1.0039
10.00	1.0000	1.0000	1.0000
20.00	0.9910	0.9938	0.9896
30.00	0.9787	0.9850	0.9760
40.00	0.9661	0.9754	0.9629
50.00	0.9564	0.9675	0.9540
60.00	0.9517	0.9627	0.9510
70.00	0.9540	0.9631	0.9553
80.00	0.9649	0.9706	0.9679
90.00	0.9839	0.9853	0.9872
100.00	1.0111	1.0085	1.0126
120.00	1.0844	1.0801	1.0759
140.00	1.1424	1.1558	1.1278
160.00	1.1048	1.1334	1.1092

Sensitivities

	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
Output Sensitivity (V/g)	2.501	2.499	2.501
Cross-axis Coupling (g/g)	0.32%	0.36%	0.52%