



REF TEK

SYSTEMS INC.

The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Topple Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651

Trimble Inc.
5217 Tennyson Parkway
Suite 400
Plano, TX 75024
Support Phone: 1-888-879-2207
Support Email: MonSol_Support@trimble.com
www.trimble.com/monitoring



REF TEK 147A-01/3/INT/SM

S/N: 16043

Normalized Magnitude-frequency Characteristics

Freq.(Hz)	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
1.00	1.0030	1.0021	1.0023
10.00	1.0000	1.0000	1.0000
20.00	0.9917	0.9941	0.9936
30.00	0.9802	0.9857	0.9844
40.00	0.9680	0.9761	0.9741
50.00	0.9581	0.9674	0.9649
60.00	0.9522	0.9610	0.9585
70.00	0.9525	0.9590	0.9569
80.00	0.9606	0.9630	0.9616
90.00	0.9765	0.9738	0.9737
100.00	1.0012	0.9929	0.9942
120.00	1.0761	1.0595	1.0638
140.00	1.1524	1.1472	1.1515
160.00	1.1236	1.1494	1.1522

Sensitivities

	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
Output Sensitivity (V/g)	2.501	2.501	2.499
Cross-axis Coupling (g/g)	0.36%	0.24%	0.60%