The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Topple Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651

Trimble Inc.
5217 Tennyson Parkway
Suite 400
Plano, TX 75024
Support Phone: 1-888-879-2207
Support Email: MonSol_Support@trimble.com
www.trimble.com/monitoring

REF TEK 147A-01/3/INT/SM
S/N: 16047

Normalized Magnitude-frequency Characteristics

Freq.(Hz)	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
1.00	1.0030	1.0027	1.0023
10.00	1.0000	1.0000	1.0000
20.00	0.9918	0.9926	0.9936
30.00	0.9802	0.9822	0.9845
40.00	0.9678	0.9708	0.9744
50.00	0.9574	0.9612	0.9656
60.00	0.9507	0.9551	0.9597
70.00	0.9500	0.9544	0.9586
80.00	0.9568	0.9610	0.9637
90.00	0.9713	0.9750	0.9755
100.00	0.9938	0.9971	0.9948
120.00	1.0592	1.0633	1.0549
140.00	1.1128	1.1244	1.1168
160.00	1.0706	1.0914	1.0965

Sensitivities

	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
Output Sensitivity (V/g)	2.500	2.499	2.499
Cross-axis Coupling (g/g)	0.20%	0.28%	0.44%