The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Topple Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651

Trimble Inc.
5217 Tennyson Parkway
Suite 400
Plano, TX 75024
Support Phone: 1-888-879-2207
Support Email: MonSol_Support@trimble.com
www.trimble.com/monitoring

REF TEK 147A-01/3/INT/SM
S/N: 16058

Normalized Magnitude-frequency Characteristics

Freq.(Hz)	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
1.00	1.0026	1.0023	1.0027
10.00	1.0000	1.0000	1.0000
20.00	0.9926	0.9936	0.9924
30.00	0.9821	0.9844	0.9816
40.00	0.9708	0.9742	0.9697
50.00	0.9611	0.9652	0.9592
60.00	0.9548	0.9590	0.9518
70.00	0.9541	0.9575	0.9497
80.00	0.9608	0.9624	0.9546
90.00	0.9753	0.9743	0.9671
100.00	0.9983	0.9943	0.9881
120.00	1.0689	1.0598	1.0571
140.00	1.1360	1.1366	1.1361
160.00	1.0967	1.1274	1.1246

Sensitivities

	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
Output Sensitivity (V/g)	2.500	2.500	2.501
Cross-axis Coupling (g/g)	0.16%	0.60%	0.60%