The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Topple Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651

Trimble Inc.
5217 Tennyson Parkway
Suite 400
Plano, TX 75024
Support Phone: 1-888-879-2207
Support Email: MonSol_Support@trimble.com
www.trimble.com/monitoring

REF TEK 147A-01/3/INT/SM
S/N: 16070

Normalized Magnitude-frequency Characteristics

Freq.(Hz)	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
1.00	1.0020	1.0023	1.0022
10.00	1.0000	1.0000	1.0000
20.00	0.9943	0.9936	0.9938
30.00	0.9863	0.9845	0.9852
40.00	0.9779	0.9746	0.9759
50.00	0.9712	0.9663	0.9683
60.00	0.9679	0.9612	0.9642
70.00	0.9698	0.9614	0.9654
80.00	0.9789	0.9687	0.9739
90.00	0.9952	0.9835	0.9898
100.00	1.0197	1.0070	1.0142
120.00	1.0914	1.0800	1.0872
140.00	1.1550	1.1552	1.1564
160.00	1.1070	1.1228	1.1167

Sensitivities

	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
Output Sensitivity (V/g)	2.501	2.500	2.499
Cross-axis Coupling (g/g)	0.44%	0.16%	0.56%