



REF TEK

SYSTEMS INC.

The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Topple Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651

Trimble Inc.
5217 Tennyson Parkway
Suite 400
Plano, TX 75024
Support Phone: 1-888-879-2207
Support Email: MonSol_Support@trimble.com
www.trimble.com/monitoring



REF TEK 147A-01/3/INT/SM

S/N: 16076

Normalized Magnitude-frequency Characteristics

Freq.(Hz)	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
1.00	1.0026	1.0026	1.0024
10.00	1.0000	1.0000	1.0000
20.00	0.9926	0.9926	0.9933
30.00	0.9822	0.9821	0.9838
40.00	0.9707	0.9703	0.9732
50.00	0.9606	0.9600	0.9640
60.00	0.9538	0.9527	0.9576
70.00	0.9524	0.9506	0.9563
80.00	0.9581	0.9555	0.9617
90.00	0.9717	0.9680	0.9744
100.00	0.9940	0.9892	0.9955
120.00	1.0655	1.0585	1.0640
140.00	1.1418	1.1362	1.1417
160.00	1.1138	1.1173	1.1270

Sensitivities

	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
Output Sensitivity (V/g)	2.501	2.501	2.501
Cross-axis Coupling (g/g)	0.24%	0.44%	0.40%