



REF TEK

SYSTEMS INC.

The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Topples Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651

Trimble Inc.
5217 Tennyson Parkway
Suite 400
Plano, TX 75024
Support Phone: 1-888-879-2207
Support Email: MonSol_Support@trimble.com
www.trimble.com/monitoring



REF TEK 147A-01/3/INT/SM

S/N: 16082

Normalized Magnitude-frequency Characteristics

Freq.(Hz)	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
1.00	1.0046	1.0047	1.0035
10.00	1.0000	1.0000	1.0000
20.00	0.9885	0.9884	0.9905
30.00	0.9752	0.9751	0.9777
40.00	0.9637	0.9637	0.9646
50.00	0.9560	0.9562	0.9544
60.00	0.9528	0.9527	0.9486
70.00	0.9548	0.9540	0.9488
80.00	0.9629	0.9606	0.9562
90.00	0.9768	0.9724	0.9706
100.00	0.9974	0.9900	0.9924
120.00	1.0597	1.0436	1.0580
140.00	1.1290	1.1070	1.1304
160.00	1.1135	1.1089	1.1207

Sensitivities

	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
Output Sensitivity (V/g)	2.500	2.499	2.502
Cross-axis Coupling (g/g)	0.24%	0.64%	0.80%