



REF TEK

SYSTEMS INC.

The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Topple Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651

Trimble Inc.
5217 Tennyson Parkway
Suite 400
Plano, TX 75024
Support Phone: 1-888-879-2207
Support Email: MonSol_Support@trimble.com
www.trimble.com/monitoring



REF TEK 147A-01/3/INT/SM

S/N: 17002

Normalized Magnitude-frequency Characteristics

Freq.(Hz)	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
1.00	1.0020	1.0022	1.0027
10.00	1.0000	1.0000	1.0000
20.00	0.9945	0.9937	0.9924
30.00	0.9868	0.9848	0.9818
40.00	0.9784	0.9750	0.9702
50.00	0.9714	0.9666	0.9605
60.00	0.9672	0.9612	0.9543
70.00	0.9676	0.9607	0.9541
80.00	0.9743	0.9669	0.9615
90.00	0.9875	0.9803	0.9773
100.00	1.0081	1.0020	1.0023
120.00	1.0710	1.0718	1.0802
140.00	1.1361	1.1524	1.1583
160.00	1.1174	1.1434	1.1187

Sensitivities

	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
Output Sensitivity (V/g)	2.501	2.500	2.501
Cross-axis Coupling (g/g)	0.04%	0.32%	0.44%