



REF TEK

SYSTEMS INC.

The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Topple Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651

Trimble Inc.
5217 Tennyson Parkway
Suite 400
Plano, TX 75024
Support Phone: 1-888-879-2207
Support Email: MonSol_Support@trimble.com
www.trimble.com/monitoring



REF TEK 147A-01/3/INT/SM

S/N: 17008

Normalized Magnitude-frequency Characteristics

Freq.(Hz)	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
1.00	1.0025	1.0022	1.0021
10.00	1.0000	1.0000	1.0000
20.00	0.9929	0.9938	0.9941
30.00	0.9830	0.9849	0.9858
40.00	0.9721	0.9750	0.9766
50.00	0.9630	0.9663	0.9688
60.00	0.9574	0.9602	0.9638
70.00	0.9573	0.9587	0.9634
80.00	0.9645	0.9635	0.9692
90.00	0.9795	0.9752	0.9816
100.00	1.0030	0.9950	1.0013
120.00	1.0747	1.0610	1.0628
140.00	1.1418	1.1420	1.1299
160.00	1.1007	1.1395	1.1221

Sensitivities

	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
Output Sensitivity (V/g)	2.499	2.500	2.500
Cross-axis Coupling (g/g)	0.20%	0.28%	0.64%