The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Topple Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651

Trimble Inc.
5217 Tennyson Parkway
Suite 400
Plano, TX 75024
Support Phone: 1-888-879-2207
Support Email: MonSol_Support@trimble.com
www.trimble.com/monitoring

REF TEK 147A-01/3/INT/SM
S/N: 17012

Normalized Magnitude-frequency Characteristics

Freq.(Hz)	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
1.00	1.0019	1.0022	1.0020
10.00	1.0000	1.0000	1.0000
20.00	0.9947	0.9938	0.9944
30.00	0.9870	0.9850	0.9863
40.00	0.9784	0.9751	0.9772
50.00	0.9708	0.9664	0.9690
60.00	0.9656	0.9603	0.9630
70.00	0.9645	0.9588	0.9611
80.00	0.9690	0.9635	0.9646
90.00	0.9798	0.9752	0.9743
100.00	0.9978	0.9948	0.9912
120.00	1.0567	1.0588	1.0477
140.00	1.1255	1.1297	1.1156
160.00	1.1170	1.1066	1.1105

Sensitivities

	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
Output Sensitivity (V/g)	2.500	2.500	2.500
Cross-axis Coupling (g/g)	0.36%	0.40%	0.32%