



REF TEK

SYSTEMS INC.

The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Tople Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651

Trimble Inc.
5217 Tennyson Parkway
Suite 400
Plano, TX 75024
Support Phone: 1-888-879-2207
Support Email: MonSol_Support@trimble.com
www.trimble.com/monitoring



REF TEK 147A-01/3/INT/SM

S/N: 17020

Normalized Magnitude-frequency Characteristics

Freq.(Hz)	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
1.00	1.0020	1.0020	1.0019
10.00	1.0000	1.0000	1.0000
20.00	0.9943	0.9945	0.9948
30.00	0.9861	0.9865	0.9872
40.00	0.9770	0.9775	0.9788
50.00	0.9690	0.9696	0.9714
60.00	0.9634	0.9641	0.9664
70.00	0.9623	0.9628	0.9655
80.00	0.9673	0.9673	0.9703
90.00	0.9792	0.9784	0.9813
100.00	0.9991	0.9970	0.9997
120.00	1.0668	1.0601	1.0605
140.00	1.1532	1.1418	1.1350
160.00	1.1535	1.1515	1.1350

Sensitivities

	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
Output Sensitivity (V/g)	2.502	2.499	2.499
Cross-axis Coupling (g/g)	0.16%	0.16%	0.52%