



REF TEK

SYSTEMS INC.

The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Topples Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651

Trimble Inc.
5217 Tennyson Parkway
Suite 400
Plano, TX 75024
Support Phone: 1-888-879-2207
Support Email: MonSol_Support@trimble.com
www.trimble.com/monitoring



REF TEK 147A-01/3/INT/SM

S/N: 17025

Normalized Magnitude-frequency Characteristics

Freq.(Hz)	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
1.00	1.0030	1.0039	1.0048
10.00	1.0000	1.0000	1.0000
20.00	0.9919	0.9897	0.9884
30.00	0.9811	0.9767	0.9752
40.00	0.9702	0.9644	0.9637
50.00	0.9619	0.9555	0.9559
60.00	0.9579	0.9513	0.9521
70.00	0.9596	0.9528	0.9534
80.00	0.9684	0.9612	0.9609
90.00	0.9841	0.9760	0.9747
100.00	1.0074	0.9981	0.9959
120.00	1.0768	1.0653	1.0635
140.00	1.1494	1.1415	1.1465
160.00	1.1274	1.1306	1.1441

Sensitivities

	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
Output Sensitivity (V/g)	2.499	2.500	2.499
Cross-axis Coupling (g/g)	0.36%	0.56%	0.12%