



REF TEK

SYSTEMS INC.

The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Tople Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651

Trimble Inc.
5217 Tennyson Parkway
Suite 400
Plano, TX 75024
Support Phone: 1-888-879-2207
Support Email: MonSol_Support@trimble.com
www.trimble.com/monitoring



REF TEK 147A-01/3/INT/SM

S/N: 17026

Normalized Magnitude-frequency Characteristics

Freq.(Hz)	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
1.00	1.0026	1.0025	1.0022
10.00	1.0000	1.0000	1.0000
20.00	0.9928	0.9931	0.9937
30.00	0.9826	0.9832	0.9848
40.00	0.9716	0.9723	0.9749
50.00	0.9624	0.9627	0.9664
60.00	0.9566	0.9562	0.9607
70.00	0.9565	0.9548	0.9599
80.00	0.9638	0.9601	0.9654
90.00	0.9788	0.9726	0.9778
100.00	1.0021	0.9934	0.9978
120.00	1.0721	1.0596	1.0609
140.00	1.1366	1.1308	1.1296
160.00	1.0990	1.1089	1.1167

Sensitivities

	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
Output Sensitivity (V/g)	2.500	2.500	2.500
Cross-axis Coupling (g/g)	0.24%	0.12%	0.56%