



REF TEK

SYSTEMS INC.

The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Topple Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651

Trimble Inc.
5217 Tennyson Parkway
Suite 400
Plano, TX 75024
Support Phone: 1-888-879-2207
Support Email: MonSol_Support@trimble.com
www.trimble.com/monitoring



REF TEK 147A-01/3/INT/SM

S/N: 17027

Normalized Magnitude-frequency Characteristics

Freq.(Hz)	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
1.00	1.0022	1.0023	1.0026
10.00	1.0000	1.0000	1.0000
20.00	0.9938	0.9934	0.9927
30.00	0.9849	0.9840	0.9822
40.00	0.9753	0.9735	0.9707
50.00	0.9672	0.9642	0.9607
60.00	0.9621	0.9576	0.9539
70.00	0.9622	0.9559	0.9526
80.00	0.9692	0.9605	0.9586
90.00	0.9836	0.9725	0.9725
100.00	1.0066	0.9928	0.9954
120.00	1.0787	1.0609	1.0705
140.00	1.1533	1.1435	1.1548
160.00	1.1178	1.1347	1.1304

Sensitivities

	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
Output Sensitivity (V/g)	2.501	2.500	2.500
Cross-axis Coupling (g/g)	0.64%	0.40%	0.40%