



REF TEK

SYSTEMS INC.

The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Tople Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651

Trimble Inc.
5217 Tennyson Parkway
Suite 400
Plano, TX 75024
Support Phone: 1-888-879-2207
Support Email: MonSol_Support@trimble.com
www.trimble.com/monitoring



REF TEK 147A-01/3/INT/SM

S/N: 17031

Normalized Magnitude-frequency Characteristics

Freq.(Hz)	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
1.00	1.0036	1.0024	1.0032
10.00	1.0000	1.0000	1.0000
20.00	0.9904	0.9934	0.9913
30.00	0.9777	0.9840	0.9796
40.00	0.9653	0.9738	0.9677
50.00	0.9563	0.9653	0.9585
60.00	0.9524	0.9601	0.9541
70.00	0.9551	0.9603	0.9562
80.00	0.9657	0.9676	0.9665
90.00	0.9833	0.9825	0.9846
100.00	1.0083	1.0058	1.0111
120.00	1.0783	1.0787	1.0874
140.00	1.1487	1.1570	1.1631
160.00	1.1362	1.1380	1.1404

Sensitivities

	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
Output Sensitivity (V/g)	2.500	2.502	2.499
Cross-axis Coupling (g/g)	0.20%	0.44%	0.56%