



REF TEK

SYSTEMS INC.

The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Topples Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651

Trimble Inc.
5217 Tennyson Parkway
Suite 400
Plano, TX 75024
Support Phone: 1-888-879-2207
Support Email: MonSol_Support@trimble.com
www.trimble.com/monitoring



REF TEK 147A-01/3/INT/SM

S/N: 17036

Normalized Magnitude-frequency Characteristics

Freq.(Hz)	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
1.00	1.0030	1.0024	1.0028
10.00	1.0000	1.0000	1.0000
20.00	0.9917	0.9933	0.9922
30.00	0.9801	0.9837	0.9810
40.00	0.9675	0.9731	0.9688
50.00	0.9570	0.9639	0.9583
60.00	0.9503	0.9578	0.9512
70.00	0.9499	0.9569	0.9497
80.00	0.9576	0.9630	0.9556
90.00	0.9737	0.9767	0.9694
100.00	0.9993	0.9991	0.9919
120.00	1.0770	1.0715	1.0631
140.00	1.1497	1.1487	1.1362
160.00	1.1054	1.1147	1.1078

Sensitivities

	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
Output Sensitivity (V/g)	2.500	2.499	2.500
Cross-axis Coupling (g/g)	0.24%	0.12%	0.40%