



REF TEK

SYSTEMS INC.

The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Tople Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651

Trimble Inc.
5217 Tennyson Parkway
Suite 400
Plano, TX 75024
Support Phone: 1-888-879-2207
Support Email: MonSol_Support@trimble.com
www.trimble.com/monitoring



REF TEK 147A-01/3/INT/SM

S/N: 17037

Normalized Magnitude-frequency Characteristics

Freq.(Hz)	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
1.00	1.0031	1.0039	1.0047
10.00	1.0000	1.0000	1.0000
20.00	0.9915	0.9897	0.9882
30.00	0.9800	0.9767	0.9746
40.00	0.9684	0.9645	0.9627
50.00	0.9597	0.9560	0.9549
60.00	0.9556	0.9523	0.9517
70.00	0.9580	0.9545	0.9536
80.00	0.9683	0.9638	0.9617
90.00	0.9859	0.9795	0.9755
100.00	1.0115	1.0024	0.9960
120.00	1.0838	1.0704	1.0575
140.00	1.1528	1.1451	1.1270
160.00	1.1258	1.1329	1.1179

Sensitivities

	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
Output Sensitivity (V/g)	2.500	2.500	2.499
Cross-axis Coupling (g/g)	0.44%	0.60%	0.24%