



REF TEK

SYSTEMS INC.

The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Tople Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651

Trimble Inc.
5217 Tennyson Parkway
Suite 400
Plano, TX 75024
Support Phone: 1-888-879-2207
Support Email: MonSol_Support@trimble.com
www.trimble.com/monitoring



REF TEK 147A-01/3/INT/SM

S/N: 17046

Normalized Magnitude-frequency Characteristics

Freq.(Hz)	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
1.00	0.9981	0.9996	0.9998
10.00	1.0000	1.0000	1.0000
20.00	1.0041	1.0010	1.0006
30.00	1.0075	1.0023	1.0013
40.00	1.0090	1.0036	1.0021
50.00	1.0086	1.0045	1.0026
60.00	1.0069	1.0047	1.0029
70.00	1.0047	1.0041	1.0029
80.00	1.0025	1.0027	1.0029
90.00	1.0013	1.0010	1.0033
100.00	1.0014	0.9995	1.0046
120.00	1.0078	1.0000	1.0123
140.00	1.0250	1.0111	1.0294
160.00	1.0479	1.0320	1.0476

Sensitivities

	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
Output Sensitivity (V/g)	2.502	2.502	2.500
Cross-axis Coupling (g/g)	0.52%	0.08%	0.28%