



REF TEK

SYSTEMS INC.

The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Tople Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651

Trimble Inc.
5217 Tennyson Parkway
Suite 400
Plano, TX 75024
Support Phone: 1-888-879-2207
Support Email: MonSol_Support@trimble.com
www.trimble.com/monitoring



REF TEK 147A-01/3/INT/SM

S/N: 17048

Normalized Magnitude-frequency Characteristics

Freq.(Hz)	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
1.00	1.0000	0.9998	0.9998
10.00	1.0000	1.0000	1.0000
20.00	0.9999	1.0004	1.0005
30.00	0.9996	1.0011	1.0012
40.00	0.9992	1.0017	1.0020
50.00	0.9987	1.0022	1.0026
60.00	0.9980	1.0025	1.0030
70.00	0.9973	1.0024	1.0030
80.00	0.9968	1.0021	1.0028
90.00	0.9969	1.0018	1.0026
100.00	0.9980	1.0018	1.0027
120.00	1.0054	1.0048	1.0064
140.00	1.0241	1.0158	1.0187
160.00	1.0518	1.0340	1.0379

Sensitivities

	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
Output Sensitivity (V/g)	2.502	2.500	2.499
Cross-axis Coupling (g/g)	0.44%	0.12%	0.24%