



REF TEK

SYSTEMS INC.

The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Topple Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651

Trimble Inc.
5217 Tennyson Parkway
Suite 400
Plano, TX 75024
Support Phone: 1-888-879-2207
Support Email: MonSol_Support@trimble.com
www.trimble.com/monitoring



REF TEK 147A-01/3/INT/SM

S/N: 17054

Normalized Magnitude-frequency Characteristics

Freq.(Hz)	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
1.00	1.0006	1.0001	1.0003
10.00	1.0000	1.0000	1.0000
20.00	0.9984	0.9997	0.9991
30.00	0.9959	0.9994	0.9978
40.00	0.9928	0.9989	0.9961
50.00	0.9894	0.9985	0.9944
60.00	0.9861	0.9982	0.9929
70.00	0.9833	0.9982	0.9918
80.00	0.9814	0.9987	0.9916
90.00	0.9810	0.9997	0.9923
100.00	0.9823	1.0015	0.9944
120.00	0.9919	1.0079	1.0033
140.00	1.0126	1.0174	1.0187
160.00	1.0376	1.0235	1.0338

Sensitivities

	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
Output Sensitivity (V/g)	2.501	2.500	2.501
Cross-axis Coupling (g/g)	0.08%	0.48%	0.08%