



REF TEK

SYSTEMS INC.

The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Topple Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651

Trimble Inc.
5217 Tennyson Parkway
Suite 400
Plano, TX 75024
Support Phone: 1-888-879-2207
Support Email: MonSol_Support@trimble.com
www.trimble.com/monitoring



REF TEK 147A-01/3/INT/SM

S/N: 17056

Normalized Magnitude-frequency Characteristics

Freq.(Hz)	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
1.00	1.0004	1.0002	1.0003
10.00	1.0000	1.0000	1.0000
20.00	0.9990	0.9994	0.9990
30.00	0.9974	0.9985	0.9974
40.00	0.9954	0.9973	0.9954
50.00	0.9934	0.9960	0.9933
60.00	0.9913	0.9947	0.9911
70.00	0.9898	0.9936	0.9894
80.00	0.9888	0.9928	0.9883
90.00	0.9889	0.9928	0.9882
100.00	0.9902	0.9938	0.9895
120.00	0.9974	1.0000	0.9973
140.00	1.0115	1.0141	1.0140
160.00	1.0268	1.0327	1.0343

Sensitivities

	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
Output Sensitivity (V/g)	2.499	2.500	2.499
Cross-axis Coupling (g/g)	0.40%	0.68%	0.24%