



REF TEK

SYSTEMS INC.

The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Topple Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651

Trimble Inc.
5217 Tennyson Parkway
Suite 400
Plano, TX 75024
Support Phone: 1-888-879-2207
Support Email: MonSol_Support@trimble.com
www.trimble.com/monitoring



REF TEK 147A-01/3/INT/SM

S/N: 17059

Normalized Magnitude-frequency Characteristics

Freq.(Hz)	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
1.00	0.9995	0.9999	0.9996
10.00	1.0000	1.0000	1.0000
20.00	1.0013	1.0002	1.0010
30.00	1.0030	1.0005	1.0025
40.00	1.0046	1.0006	1.0039
50.00	1.0056	1.0005	1.0049
60.00	1.0057	1.0001	1.0052
70.00	1.0048	0.9993	1.0046
80.00	1.0032	0.9985	1.0031
90.00	1.0014	0.9979	1.0013
100.00	1.0000	0.9980	0.9995
120.00	1.0016	1.0026	0.9994
140.00	1.0137	1.0168	1.0094
160.00	1.0340	1.0375	1.0284

Sensitivities

	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
Output Sensitivity (V/g)	2.501	2.499	2.501
Cross-axis Coupling (g/g)	0.40%	0.20%	0.16%