



REF TEK

SYSTEMS INC.

The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Tople Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651

Trimble Inc.
5217 Tennyson Parkway
Suite 400
Plano, TX 75024
Support Phone: 1-888-879-2207
Support Email: MonSol_Support@trimble.com
www.trimble.com/monitoring



REF TEK 147A-01/3/INT/SM

S/N: 17061

Normalized Magnitude-frequency Characteristics

Freq.(Hz)	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
1.00	1.0002	1.0001	1.0000
10.00	1.0000	1.0000	1.0000
20.00	0.9995	0.9996	1.0000
30.00	0.9986	0.9989	0.9999
40.00	0.9976	0.9981	0.9997
50.00	0.9964	0.9972	0.9993
60.00	0.9953	0.9963	0.9986
70.00	0.9945	0.9957	0.9978
80.00	0.9941	0.9956	0.9968
90.00	0.9944	0.9961	0.9961
100.00	0.9958	0.9977	0.9958
120.00	1.0029	1.0051	0.9983
140.00	1.0178	1.0203	1.0094
160.00	1.0370	1.0395	1.0301

Sensitivities

	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
Output Sensitivity (V/g)	2.500	2.500	2.502
Cross-axis Coupling (g/g)	0.44%	0.28%	0.16%