



REF TEK

SYSTEMS INC.

The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Tople Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651

Trimble Inc.
5217 Tennyson Parkway
Suite 400
Plano, TX 75024
Support Phone: 1-888-879-2207
Support Email: MonSol_Support@trimble.com
www.trimble.com/monitoring



REF TEK 147A-01/3/INT/SM

S/N: 17063

Normalized Magnitude-frequency Characteristics

Freq.(Hz)	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
1.00	1.0005	1.0002	1.0004
10.00	1.0000	1.0000	1.0000
20.00	0.9986	0.9993	0.9988
30.00	0.9964	0.9982	0.9970
40.00	0.9936	0.9968	0.9949
50.00	0.9908	0.9952	0.9927
60.00	0.9881	0.9937	0.9908
70.00	0.9861	0.9925	0.9897
80.00	0.9850	0.9919	0.9895
90.00	0.9853	0.9921	0.9907
100.00	0.9872	0.9934	0.9935
120.00	0.9971	1.0010	1.0048
140.00	1.0154	1.0170	1.0236
160.00	1.0345	1.0363	1.0419

Sensitivities

	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
Output Sensitivity (V/g)	2.501	2.501	2.502
Cross-axis Coupling (g/g)	0.48%	0.12%	0.44%