The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Topple Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651

Trimble Inc.
5217 Tennyson Parkway
Suite 400
Plano, TX 75024
Support Phone: 1-888-879-2207
Support Email: MonSol_Support@trimble.com
www.trimble.com/monitoring

REF TEK 147A-01/3/INT/SM
S/N: 17064

Normalized Magnitude-frequency Characteristics

Freq.(Hz)	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
1.00	1.0000	1.0001	1.0004
10.00	1.0000	1.0000	1.0000
20.00	1.0000	0.9997	0.9989
30.00	0.9999	0.9993	0.9973
40.00	0.9998	0.9989	0.9954
50.00	0.9997	0.9986	0.9935
60.00	0.9996	0.9985	0.9919
70.00	0.9997	0.9989	0.9911
80.00	0.9999	1.0000	0.9912
90.00	1.0004	1.0018	0.9926
100.00	1.0014	1.0045	0.9955
120.00	1.0059	1.0138	1.0065
140.00	1.0153	1.0284	1.0241
160.00	1.0276	1.0433	1.0408

Sensitivities

	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
Output Sensitivity (V/g)	2.500	2.499	2.499
Cross-axis Coupling (g/g)	0.56%	0.56%	0.20%