



REF TEK

SYSTEMS INC.

The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Topple Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651

Trimble Inc.
5217 Tennyson Parkway
Suite 400
Plano, TX 75024
Support Phone: 1-888-879-2207
Support Email: MonSol_Support@trimble.com
www.trimble.com/monitoring



REF TEK 147A-01/3/INT/SM

S/N: 17065

Normalized Magnitude-frequency Characteristics

Freq.(Hz)	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
1.00	1.0000	1.0001	0.9998
10.00	1.0000	1.0000	1.0000
20.00	0.9999	0.9998	1.0007
30.00	0.9997	0.9994	1.0017
40.00	0.9995	0.9990	1.0030
50.00	0.9994	0.9986	1.0042
60.00	0.9995	0.9982	1.0052
70.00	0.9998	0.9979	1.0060
80.00	1.0005	0.9980	1.0064
90.00	1.0018	0.9986	1.0066
100.00	1.0039	0.9999	1.0067
120.00	1.0112	1.0057	1.0085
140.00	1.0226	1.0179	1.0160
160.00	1.0320	1.0346	1.0304

Sensitivities

	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
Output Sensitivity (V/g)	2.502	2.499	2.501
Cross-axis Coupling (g/g)	0.12%	0.60%	0.16%