The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Topple Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651

Trimble Inc.
5217 Tennyson Parkway
Suite 400
Plano, TX 75024
Support Phone: 1-888-879-2207
Support Email: MonSol_Support@trimble.com
www.trimble.com/monitoring

REF TEK 147A-01/3/INT/SM
S/N: 17068

Normalized Magnitude-frequency Characteristics

Freq.(Hz)	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
1.00	1.0002	1.0004	1.0001
10.00	1.0000	1.0000	1.0000
20.00	0.9994	0.9989	0.9997
30.00	0.9984	0.9973	0.9991
40.00	0.9973	0.9954	0.9985
50.00	0.9961	0.9934	0.9978
60.00	0.9952	0.9916	0.9973
70.00	0.9947	0.9905	0.9970
80.00	0.9950	0.9904	0.9971
90.00	0.9964	0.9917	0.9979
100.00	0.9990	0.9945	0.9994
120.00	1.0094	1.0064	1.0057
140.00	1.0279	1.0271	1.0171
160.00	1.0485	1.0478	1.0293

Sensitivities

	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
Output Sensitivity (V/g)	2.500	2.499	2.500
Cross-axis Coupling (g/g)	0.04%	0.16%	0.52%