



REF TEK

SYSTEMS INC.

The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Topple Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651

Trimble Inc.
5217 Tennyson Parkway
Suite 400
Plano, TX 75024
Support Phone: 1-888-879-2207
Support Email: MonSol_Support@trimble.com
www.trimble.com/monitoring



REF TEK 147A-01/3/INT/SM

S/N: 17072

Normalized Magnitude-frequency Characteristics

Freq.(Hz)	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
1.00	1.0002	1.0001	1.0000
10.00	1.0000	1.0000	1.0000
20.00	0.9993	0.9998	0.9999
30.00	0.9982	0.9995	0.9996
40.00	0.9968	0.9992	0.9993
50.00	0.9953	0.9989	0.9988
60.00	0.9939	0.9988	0.9982
70.00	0.9928	0.9990	0.9975
80.00	0.9923	0.9997	0.9968
90.00	0.9927	1.0011	0.9963
100.00	0.9942	1.0033	0.9961
120.00	1.0022	1.0113	0.9978
140.00	1.0185	1.0249	1.0041
160.00	1.0380	1.0394	1.0145

Sensitivities

	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
Output Sensitivity (V/g)	2.502	2.499	2.501
Cross-axis Coupling (g/g)	0.08%	0.04%	0.08%