



REF TEK

SYSTEMS INC.

The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Topple Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651

Trimble Inc.
5217 Tennyson Parkway
Suite 400
Plano, TX 75024
Support Phone: 1-888-879-2207
Support Email: MonSol_Support@trimble.com
www.trimble.com/monitoring



REF TEK 147A-01/3/INT/SM

S/N: 17074

Normalized Magnitude-frequency Characteristics

Freq.(Hz)	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
1.00	0.9998	0.9997	0.9998
10.00	1.0000	1.0000	1.0000
20.00	1.0006	1.0008	1.0006
30.00	1.0014	1.0020	1.0015
40.00	1.0025	1.0036	1.0026
50.00	1.0038	1.0055	1.0038
60.00	1.0052	1.0078	1.0050
70.00	1.0067	1.0102	1.0060
80.00	1.0085	1.0130	1.0071
90.00	1.0106	1.0161	1.0080
100.00	1.0132	1.0197	1.0091
120.00	1.0213	1.0289	1.0126
140.00	1.0348	1.0407	1.0200
160.00	1.0485	1.0484	1.0302

Sensitivities

	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
Output Sensitivity (V/g)	2.500	2.499	2.500
Cross-axis Coupling (g/g)	0.32%	0.04%	0.20%