



REF TEK

SYSTEMS INC.

The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Topple Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651

Trimble Inc.
5217 Tennyson Parkway
Suite 400
Plano, TX 75024
Support Phone: 1-888-879-2207
Support Email: MonSol_Support@trimble.com
www.trimble.com/monitoring



REF TEK 147A-01/3/INT/SM

S/N: 17076

Normalized Magnitude-frequency Characteristics

Freq.(Hz)	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
1.00	1.0002	1.0007	0.9999
10.00	1.0000	1.0000	1.0000
20.00	0.9994	0.9981	1.0003
30.00	0.9986	0.9953	1.0008
40.00	0.9978	0.9919	1.0014
50.00	0.9972	0.9885	1.0021
60.00	0.9971	0.9855	1.0028
70.00	0.9978	0.9835	1.0034
80.00	0.9996	0.9829	1.0040
90.00	1.0025	0.9841	1.0047
100.00	1.0067	0.9872	1.0057
120.00	1.0195	0.9997	1.0091
140.00	1.0362	1.0188	1.0163
160.00	1.0478	1.0337	1.0257

Sensitivities

	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
Output Sensitivity (V/g)	2.501	2.500	2.501
Cross-axis Coupling (g/g)	0.72%	0.08%	0.64%