



REF TEK

SYSTEMS INC.

The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Topple Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651

Trimble Inc.
5217 Tennyson Parkway
Suite 400
Plano, TX 75024
Support Phone: 1-888-879-2207
Support Email: MonSol_Support@trimble.com
www.trimble.com/monitoring



REF TEK 147A-01/3/INT/SM

S/N: 17078

Normalized Magnitude-frequency Characteristics

Freq.(Hz)	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
1.00	1.0001	1.0003	0.9999
10.00	1.0000	1.0000	1.0000
20.00	0.9998	0.9991	1.0004
30.00	0.9994	0.9977	1.0009
40.00	0.9988	0.9960	1.0016
50.00	0.9980	0.9942	1.0022
60.00	0.9971	0.9925	1.0028
70.00	0.9962	0.9913	1.0033
80.00	0.9954	0.9908	1.0040
90.00	0.9951	0.9914	1.0049
100.00	0.9955	0.9935	1.0064
120.00	1.0002	1.0034	1.0133
140.00	1.0129	1.0227	1.0290
160.00	1.0306	1.0446	1.0514

Sensitivities

	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
Output Sensitivity (V/g)	2.500	2.500	2.500
Cross-axis Coupling (g/g)	0.32%	0.40%	0.12%