The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Topple Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651

Trimble Inc.
5217 Tennyson Parkway
Suite 400
Plano, TX 75024
Support Phone: 1-888-879-2207
Support Email: MonSol_Support@trimble.com
www.trimble.com/monitoring

REF TEK 147A-01/3/INT/SM
S/N: 18103

Normalized Magnitude-frequency Characteristics

Freq.(Hz)	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
1.00	1.0005	0.9999	1.0006
10.00	1.0000	1.0000	1.0000
20.00	0.9987	1.0003	0.9982
30.00	0.9967	1.0007	0.9956
40.00	0.9944	1.0012	0.9926
50.00	0.9923	1.0017	0.9898
60.00	0.9905	1.0023	0.9877
70.00	0.9898	1.0029	0.9869
80.00	0.9904	1.0038	0.9879
90.00	0.9926	1.0051	0.9909
100.00	0.9966	1.0071	0.9961
120.00	1.0108	1.0151	1.0137
140.00	1.0315	1.0319	1.0378
160.00	1.0473	1.0542	1.0547

Sensitivities

	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
Output Sensitivity (V/g)	2.501	2.500	2.500
Cross-axis Coupling (g/g)	0.32%	0.08%	0.04%