



REF TEK

SYSTEMS INC.

The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Tople Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651

Trimble Inc.
5217 Tennyson Parkway
Suite 400
Plano, TX 75024
Support Phone: 1-888-879-2207
Support Email: MonSol_Support@trimble.com
www.trimble.com/monitoring



REF TEK 147A-01/3/INT/SM

S/N: 18104

Normalized Magnitude-frequency Characteristics

Freq.(Hz)	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
1.00	0.9998	1.0003	0.9999
10.00	1.0000	1.0000	1.0000
20.00	1.0006	0.9992	1.0003
30.00	1.0016	0.9980	1.0008
40.00	1.0027	0.9966	1.0014
50.00	1.0039	0.9956	1.0021
60.00	1.0051	0.9950	1.0029
70.00	1.0061	0.9954	1.0036
80.00	1.0070	0.9970	1.0046
90.00	1.0077	1.0000	1.0056
100.00	1.0085	1.0045	1.0070
120.00	1.0113	1.0185	1.0118
140.00	1.0188	1.0370	1.0210
160.00	1.0306	1.0495	1.0326

Sensitivities

	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
Output Sensitivity (V/g)	2.500	2.502	2.501
Cross-axis Coupling (g/g)	0.44%	0.04%	0.00%