



REF TEK

SYSTEMS INC.

The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Tople Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651

Trimble Inc.
5217 Tennyson Parkway
Suite 400
Plano, TX 75024
Support Phone: 1-888-879-2207
Support Email: MonSol_Support@trimble.com
www.trimble.com/monitoring



REF TEK 147A-01/3/INT/SM

S/N: 18106

Normalized Magnitude-frequency Characteristics

Freq.(Hz)	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
1.00	1.0006	1.0003	1.0006
10.00	1.0000	1.0000	1.0000
20.00	0.9984	0.9992	0.9984
30.00	0.9959	0.9980	0.9959
40.00	0.9931	0.9967	0.9931
50.00	0.9905	0.9954	0.9904
60.00	0.9884	0.9944	0.9883
70.00	0.9873	0.9940	0.9873
80.00	0.9877	0.9945	0.9877
90.00	0.9896	0.9961	0.9899
100.00	0.9931	0.9989	0.9938
120.00	1.0044	1.0088	1.0068
140.00	1.0170	1.0238	1.0242
160.00	1.0205	1.0364	1.0365

Sensitivities

	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
Output Sensitivity (V/g)	2.501	2.502	2.501
Cross-axis Coupling (g/g)	0.04%	0.36%	0.68%