



REF TEK

SYSTEMS INC.

The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Topple Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651

Trimble Inc.
5217 Tennyson Parkway
Suite 400
Plano, TX 75024
Support Phone: 1-888-879-2207
Support Email: MonSol_Support@trimble.com
www.trimble.com/monitoring



REF TEK 147A-01/3/INT/SM

S/N: 18108

Normalized Magnitude-frequency Characteristics

Freq.(Hz)	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
1.00	0.9999	1.0002	0.9999
10.00	1.0000	1.0000	1.0000
20.00	1.0003	0.9993	1.0002
30.00	1.0009	0.9982	1.0005
40.00	1.0015	0.9969	1.0010
50.00	1.0021	0.9956	1.0015
60.00	1.0026	0.9945	1.0021
70.00	1.0029	0.9939	1.0028
80.00	1.0031	0.9941	1.0038
90.00	1.0032	0.9953	1.0050
100.00	1.0031	0.9978	1.0067
120.00	1.0033	1.0080	1.0124
140.00	1.0051	1.0255	1.0229
160.00	1.0079	1.0435	1.0357

Sensitivities

	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
Output Sensitivity (V/g)	2.499	2.500	2.502
Cross-axis Coupling (g/g)	0.44%	0.00%	0.04%