The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Topple Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651

Trimble Inc.
5217 Tennyson Parkway
Suite 400
Plano, TX 75024
Support Phone: 1-888-879-2207
Support Email: MonSol_Support@trimble.com
www.trimble.com/monitoring

REF TEK 147A-01/3/INT/SM
S/N: 18112

Normalized Magnitude-frequency Characteristics

Freq.(Hz)	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
1.00	1.0004	0.9995	1.0002
10.00	1.0000	1.0000	1.0000
20.00	0.9987	1.0013	0.9993
30.00	0.9968	1.0032	0.9982
40.00	0.9944	1.0050	0.9969
50.00	0.9920	1.0064	0.9956
60.00	0.9897	1.0070	0.9945
70.00	0.9881	1.0066	0.9939
80.00	0.9874	1.0054	0.9941
90.00	0.9880	1.0040	0.9954
100.00	0.9900	1.0031	0.9980
120.00	0.9993	1.0065	1.0081
140.00	1.0161	1.0237	1.0248
160.00	1.0338	1.0518	1.0402

Sensitivities

	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
Output Sensitivity (V/g)	2.501	2.500	2.502
Cross-axis Coupling (g/g)	0.12%	0.08%	0.00%