



REF TEK

SYSTEMS INC.

The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Topple Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651

Trimble Inc.
5217 Tennyson Parkway
Suite 400
Plano, TX 75024
Support Phone: 1-888-879-2207
Support Email: MonSol_Support@trimble.com
www.trimble.com/monitoring



REF TEK 147A-01/3/INT/SM

S/N: 18114

Normalized Magnitude-frequency Characteristics

Freq.(Hz)	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
1.00	1.0005	1.0002	1.0005
10.00	1.0000	1.0000	1.0000
20.00	0.9986	0.9994	0.9986
30.00	0.9966	0.9986	0.9966
40.00	0.9941	0.9975	0.9941
50.00	0.9915	0.9964	0.9914
60.00	0.9890	0.9954	0.9890
70.00	0.9873	0.9948	0.9872
80.00	0.9866	0.9946	0.9864
90.00	0.9873	0.9953	0.9869
100.00	0.9898	0.9971	0.9890
120.00	1.0011	1.0051	0.9991
140.00	1.0216	1.0207	1.0166
160.00	1.0424	1.0389	1.0326

Sensitivities

	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
Output Sensitivity (V/g)	2.499	2.501	2.502
Cross-axis Coupling (g/g)	0.12%	0.36%	0.36%