



REF TEK

SYSTEMS INC.

The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Topple Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651

Trimble Inc.
5217 Tennyson Parkway
Suite 400
Plano, TX 75024
Support Phone: 1-888-879-2207
Support Email: MonSol_Support@trimble.com
www.trimble.com/monitoring



REF TEK 147A-01/3/INT/SM

S/N: 18119

Normalized Magnitude-frequency Characteristics

Freq.(Hz)	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
1.00	1.0006	1.0003	1.0006
10.00	1.0000	1.0000	1.0000
20.00	0.9981	0.9993	0.9983
30.00	0.9953	0.9981	0.9957
40.00	0.9919	0.9968	0.9927
50.00	0.9884	0.9954	0.9900
60.00	0.9852	0.9942	0.9879
70.00	0.9828	0.9935	0.9872
80.00	0.9816	0.9935	0.9883
90.00	0.9819	0.9946	0.9915
100.00	0.9839	0.9970	0.9970
120.00	0.9929	1.0066	1.0150
140.00	1.0071	1.0236	1.0385
160.00	1.0182	1.0410	1.0519

Sensitivities

	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
Output Sensitivity (V/g)	2.499	2.499	2.499
Cross-axis Coupling (g/g)	0.12%	0.56%	0.00%