The calibration sheet attached is the original calibration sheet for your REF TEK product and includes contact information and branding for the company that performed the testing.

Reftek Systems Inc. purchased the REF TEK product line in May 2020.

Reftek Systems will honour existing product warranties and extended warranties.

ANY QUESTIONS OR REQUESTS FOR SUPPORT SHOULD BE ADDRESSED AS FOLLOWS:

Sales & Administrative: sales@reftek.com

Support & Engineering: support@reftek.com

Accounting: accounting@reftek.com

Address: 36 Topple Drive, Dartmouth, NS Canada B3B 1L6

Phone: +1-902-444-7650

Fax: +1-902-444-7651

Trimble Inc.
5217 Tennyson Parkway
Suite 400
Plano, TX 75024
Support Phone: 1-888-879-2207
Support Email: MonSol_Support@trimble.com
www.trimble.com/monitoring

REF TEK 147A-01/3/INT/SM
S/N: 18123

Normalized Magnitude-frequency Characteristics

Freq.(Hz)	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
1.00	1.0001	1.0002	1.0001
10.00	1.0000	1.0000	1.0000
20.00	0.9996	0.9995	0.9996
30.00	0.9991	0.9988	0.9990
40.00	0.9985	0.9979	0.9983
50.00	0.9978	0.9969	0.9974
60.00	0.9973	0.9961	0.9966
70.00	0.9971	0.9957	0.9959
80.00	0.9973	0.9959	0.9955
90.00	0.9983	0.9970	0.9956
100.00	1.0001	0.9991	0.9965
120.00	1.0074	1.0082	1.0013
140.00	1.0203	1.0254	1.0111
160.00	1.0337	1.0464	1.0216

Sensitivities

	Z-Axis (Vertical)	Y-Axis (North/South)	X-Axis (East/West)
Output Sensitivity (V/g)	2.500	2.502	2.502
Cross-axis Coupling (g/g)	0.28%	0.24%	0.04%